









## **VOLTAGE CHANGE PROCEDURE FOR CP-45**

- Disconnect unit from the AC line and the system.
- Remove the top cover
- Locate the power supply section: Varistor RV1 and jumpers J3, J4 and J5
- For 100-120VAC, jumpers have to be installed in J3 and J5, jumper J4 has to be removed.
- Install varistor with the proper value for 100-120VAC operation.
- Fuse remains the same.
- For 220-240VAC, jumper J4 has to be installed, jumpers J3 and J5 have to be removed.
- Install varistor with the proper value for 220-240VAC operation.
- Fuse remains the same.
- Test unit for proper operation and re-install top cover.